

# GN300 Datasheet



AQL 1.5

EN455 Parts 1, 2, 3 & 4

MDD 93/42/EEC

EN374

Class 1 Medical Device



Stretch Nitrile Blue Powder Free Gloves

## General Details

Material	Stretch Nitrile
Surface Finish	Powder Free (inner chlorinated., outer polymer coated)
Colour	Blue
Geometry	Ambidextrous
Texture	Fully Textured (Palm Textured)
Cuff and Finishing	Beaded

## Glove Specifications

Size	Weight g/pc	Length mm	Palm Width mm	Cuff	Thickness Palm	Finger
XS	2.7	240	≤ 80	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
S	2.8	240	85 ± 5	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
M	3.2	240	95 ± 5	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
L	3.7	240	105 ± 5	0.095 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
XL	4.2	240	≥ 110	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
Powder Content (Maximum)			2 mg / glove			
Protein Content (Maximum)			Not applicable			

## Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

## Manufacturing Standards

Manufacturing Standards	Incompliance with ISO 9001:2008
Presentation	Standard: 200 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only.



**G·S·L**

WWW.GLOVEMAN.CO.UK

**Gloveman Supplies Ltd**

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 310734

f. 01209 314760